

Performance of the AugerPrime Surface Detector

2 November 2021

Alexander Streich


Pierre Auger Observatory



Fluorescence Detector (FD)
with 27 telescopes




Surface Detector (SD)
with 1660 stations



AugerPrime – Upgrade

Radio antenna (RD)



Scintillation detector (SSD)


Underground detector (UMD)

Electronics board (UUB)

Small photomultiplier tube (SPMT)


Date: 01/11/2021
SSD deployed: 1377
• SD station

■ UUB pre-prod.: 79
● SSD pre-prod.: 76
▲ SSD w/o PMT: 1221



Upgraded Unified Board (UUB)

- Search for electrical transients
- Noise levels of input channels
- Amplification factors of input channels




Pre-production version (2021)




Upgraded Unified Board (UUB)

- Noise levels of input channels:

Prototype V1 (2016)



Pre-production version (2021)





→ Similar noise levels for original UB

Upgraded Unified Board (UUB)

- Amplification factors of input channels:

$$\text{Signal: } S = \sum_{t_{\text{start}}}^{t_{\text{stop}}} (s(t)) - N\bar{b}$$



$$\text{Amplification factor: } A = \frac{S_{\text{HG}}}{S_{\text{LG}}}$$



Surface Scintillator Detector (SSD)


- Light tightness tests:

$$\text{Skewness: } Sk = \frac{\frac{1}{N} \sum (b(t) - b)^3}{\sigma_b^3}$$




Calibration Procedure

- Optimization of algorithm + adaptations for AugerPrime stations



Calibration Procedure

- Failing current fit procedure:





Calibration Procedure

- Modified fit procedure:


C_i

$$d_i = \frac{1}{3} \left(\frac{C_{i-1}}{\Delta_{i-1}} + \frac{C_i}{\Delta_i} + \frac{C_{i+1}}{\Delta_{i+1}} \right)$$



Calibration Procedure

- Successful modified fit procedure:





August 2021:

Current: 94.83 %

New: 99.66 %



Calibration Procedure

- Comparison current to modified (data from August 2021):




Calibration Procedure

- Comparison online estimate to modified:



Calibration Procedure

- MIP charge stability:



Summary

- Electronics board properties match requirements ✓
- **1377** out of **1518** SSD units deployed: stable long-term performance ✓
- Single SSD with mechanical failure
- Modified calibration procedure with improved stability and flexibility ✓

Thank You!

Backup

- Rejected PMTs:

Non-upgraded UB stations

